

KROFtools®

Inspection Certificate

Model	3/4" 800N.m	Serial Number		K221100055
Style	Mechanical prefabricated	Calibration Instrument		PB6036
Range	200-800N.M	Instrument Precision		0.50%
Fixed Text Value	Clockwise Direction Torque Value			Max Error%
200	203.40	203.30	203.20	1.65
480	483.30	485.00	485.40	0.95
800	821.30	818.00	818.10	2.39

Test Result: The test torque value meets the standard of 1S06789, GB/T15729 and ASME B107.14M.

We recommend you to verify this torque wrench after working for 12 months or 5000 times