

KROFtools®

Inspection Certificate

Model	1" 1500N.m	Serial Number		K221100028
Style	Mechanical prefabricated	Calibration Instrument		PB6036
Range	300-1500N.M	Instrument Precision		0.50%
Fixed Text Value	Clockwise Direction Torque Value			Max Error%
300	308.20	307.30	307.10	2.51
900	897.30	897.90	897.50	-0.27
1500	1501.70	1495.40	1483.10	-0.44

Test Result: The test torque value meets the standard of 1506789, GB/T15729 and ASME B107.14M.

We recommend you to verify this torque wrench after working for 12 months or 5000 times